

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p><b>22215</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Inivos Limited, trading as Test Labs</b></p> <p><b>Issue No: 002      Issue date: 08 August 2022</b></p>	
	<p><b>Unit 22 Coningsby Road</b> <b>Bretton</b> <b>Peterborough</b> <b>PE3 8SB</b></p>	<p><b>Contact: Toni Carlton</b> <b>Tel: +44 (0)8452706690</b> <b>E-Mail: <a href="mailto:toni.carlton@testlabsuk.com">toni.carlton@testlabsuk.com</a></b> <b>Website: <a href="http://www.testlabsuk.com">www.testlabsuk.com</a></b></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
CHEMICAL DISINFECTANTS Including surface disinfectants, handwash, hand rubs	<p><u>Microbiological Quantitative Disinfectant Suspension Testing:</u></p> <p>Bacterial quantitative suspension test for food, industrial, domestic and institutional applications</p> <p>Fungicidal or yeasticidal quantitative suspension test for food, industrial, domestic and institutional applications</p> <p>Bactericidal quantitative suspension test for medical applications</p> <p>Fungicidal or yeasticidal quantitative suspension test for medical applications</p>	<p>Documented in-house methods based on BS EN standard methods, including customer specified conditions</p> <p>MIC-TP-001 based on BS EN 1276:2019</p> <p>MIC-TP-003 based on BS EN 1650:2019</p> <p>MIC-TP-002 based on BS EN 13727:2012 +A2:2015</p> <p>MIC-TP-005 based on BS EN 13624:2013</p>
Impregnated wipes	<p><u>Microbiological Quantitative Non-Porous Surface Disinfectant Testing:</u></p> <p>Bactericidal and fungicidal or yeasticidal quantitative non-porous surface test for medical applications (mechanical action employing wipes)</p>	<p>MIC-TP-007 based on BS EN 16615:2015</p>
AUTOMATED ROOM DISINFECTION MACHINES	<p>Bactericidal, sporicidal, fungicidal and yeasticidal evaluation of automated airborne disinfection process</p>	<p>MIC-TP-004 based on BS EN 17272:2020 employing test chamber sizes of 67.5m<sup>3</sup> or 11.3m<sup>3</sup></p>
<p>END</p>		